Medtronic



Effective protection. Effective humidification.


DAR[™] filter HMEs reach peak performance.

The greatest defense against pathogen harm is to prevent it from happening in the first place. DAR™ filter heat and moisture exchangers (HMEs) ensure effective airway humidification in patients to lower the incidence of infections.

The DAR $^{\text{TM}}$ filter HME first captures heat and water vapor from a patient's exhaled air. It then adds that heat and moisture to the patient's inspired air to provide optimal humidification.

That process delivers an advanced level of humidification performance. Three DAR™ filter HMEs ranked among the top 10 for humidity output¹ in a recently published study of 48 filters and HMEs.

Measured absolute humidity in independent published testing



Source: Lellouche et al



Mechanical filter HME, large



Electrostatic filter HME, small, angled port



Electrostatic filter HME, large



Pediatric electrostatic filter HME, small



Electrostatic filter HME,



Infant-pediatric electrostatic filter, small

ELECTROSTATIC FILTER HMEs					
	Large	Small	Small, Angled Port	Pediatric	Infant-Pediatric
Catalog number	352U5805	352U5877	352U5996	355U5430	355U5427
Quantity/box	50	50	50	50	50
Recommended tidal volume	300-1500 mL	150-1200 mL	150-1200 mL	75-300 mL	30-100 mL
Moisture output					
Vt 50 mL					28 mg H ₂ O/L ²
Vt 250 mL	33.9 mg H ₂ O/L ²	34.4 mg H ₂ O/L ⁴	34.4 mg H ₂ O/L ⁴	31 mg H ₂ O/L ²	
Vt 500 mL	33.3 mg H ₂ O/L ²	33.6 mg H ₂ O/L ²	33.6 mg H ₂ O/L ²		
Vt 1000 mL	32.4 mg H ₂ O/L ²	32.9 mg H ₂ O/L ⁴	32.9 mg H ₂ O/L ⁴		
Moisture loss*	6 mg H ₂ O/L at Vt 500 mL	6 mg H ₂ O/L at Vt 500 mL ⁴	6 mg H ₂ O/L at Vt 500 mL	6 mg H ₂ O/L at Vt 75 mL	NA
Resistance to flow before use	(ISO 9360)				
5 L/min					0.6 cm H ₂ O
15 L/min				1.4 cm H ₂ O	2.5 cm H ₂ O
30 L/min	1.0 cm H ₂ O	1.2 cm H ₂ O	1.2 cm H ₂ O	3.0 cm H ₂ O	
60 L/min	2.1 cm H ₂ O	2.8 cm H ₂ O	2.9 cm H ₂ O		
90 L/min	3.7 cm H ₂ O	4.8 cm H ₂ O	5.2 cm H ₂ O		
Filtration efficiency					
Bacterial	>99.9999%	>99.9998%	>99.9998%	>99.999%	>99.999%
Viral	>99.998%	>99.999%	>99.999%	>99.99%	>99.99%
NaCl3	>99.623%	>98.352%6	>98.352%6	>96.263%	>94.186%
Internal volume	93 mL	51 mL	61 mL	29 mL	10 mL
Weight (approx.)	48 g	28 g	29 g	21 g	9 g
Type of filtration	Electrostatic	Electrostatic	Electrostatic	Electrostatic	Electrostatic

 $\mathsf{DAR}^{\scriptscriptstyle\mathsf{TM}}$ filter HMEs have been tested against microbes as small as 0.02 $\mu.$

MECHANICAL FILTER HMEs		
	Large	
Catalog number	354U5876	
Quantity/box	50	
Recommended tidal volume	300-1500 mL	
Moisture output		
Vt 50 mL		
Vt 250 mL	34.7 mg H ₂ O/L ⁵	
Vt 500 mL	34.1 mg H ₂ O/L ²	
Vt 1000 mL	33.4 mg H ₂ O/L ⁵	
Moisture loss*	5 mg H ₂ O/L at Vt 500 mL	
Resistance to flow before use (ISO	9360)	
5 L/min		
15 L/min		
30 L/min	1.1 cm H ₂ O	
60 L/min	2.5 cm H ₂ O	
90 L/min	4.2 cm H ₂ O	
Filtration efficiency		
Bacterial	≥99.9999%	
Viral	≥99.9999%	
NaCl ⁴	≥99.764%	
Internal volume	96 mL	
Weight (approx.)	49 g	
Type of filtration	Mechanical	

DAR $^{\!\scriptscriptstyle{\text{TM}}}$ filter HMEs have been tested against microbes as small as 0.02 μ .

*Internal testing Mirandola (various 2005-2008).

©2021 Medtronic. Medtronic, Medtronic logo, and Engineering the extraordinary are trademarks of Medtronic. All other brands are trademarks of a Medtronic company. 12/2021-US-RE-2100633-[WF#3593979]

6135 Gunbarrel Avenue Boulder, CO 80301 800.635.5267 medtronic.com/covidien



^{1.} Lellouche F, Taillé S, Lefrançois F, et al. Humidification performance of 48 passive airway humidifiers: comparison with manufacturer data. *Chest.* 2009;135(2):276-286.

 $^{2.\} MHRA.\ Evaluation\ no.\ 04005: Breathing\ system\ filters, an\ assessment\ of\ 104\ breathing\ system\ filters.\ March\ 2004.$

^{3.} Nelson Laboratories Inc. Sodium chloride aerosol testing of breathing system filters (BSF). Lab.No. 399951A.1 Amended. January 2008.

^{4.} TIM, Technologie-Institut Medizin GmbH - Universitätsklinikum Göttingen, Germany. HME-Test Report 2008/22 DAR Hygrobac "S". July 2008.

TIM, Technologie-Institut Medizin GmbH - Universitätsklinikum Göttingen, Germany. HME-Test. Report 2009/04 DAR Hygroster. May 2009.

Nelson Laboratories Inc. Sodium chloride aerosol testing of breathing system filters (BSF). Lab.No. 717597. November 2013.